

# CERTIFICATE OF ANALYSIS

CANNON® CERTIFIED VISCOSITY REFERENCE STANDARD					
Viscosity Standard: <b>N10</b>			Lot Number: <b>16301</b>		
Certification/Issue Date: <b>01/27/2017</b>			Expiry Date: <b>01/31/2019</b>		
Temperature		Kinematic Viscosity	Dynamic Viscosity	Density	
° C	° F	mm <sup>2</sup> /s (cSt)	mPa·s (cP)	g/cm <sup>3</sup> (g/mL)	
20.00	68.00	21.43	18.62	0.8691	
25.00	77.00	17.26	14.94	0.8658	
37.78	100.00	10.65	9.131	0.8574	
40.00	104.00	9.883	8.459	0.8559	
50.00	122.00	7.263	6.169	0.8493	
80.00	176.00	3.557	2.950	0.8295	
98.89	210.00	2.540	2.075	0.8170	
100.00	212.00	2.495	2.036	0.8163	
Tested and certified in the U.S.A.					

This Certificate of Analysis shall not be reproduced, except in full, without the written approval of CANNON Instrument Company.



### USAGE INFORMATION<sup>1</sup>

**Intended Use and Instructions:** This CANNON® Certified Viscosity Reference Standard is intended for but not restricted to the calibration and performance verification of various types of viscometers or density measurement equipment. Consult user's manual and test methods specific to your equipment for operating instructions and procedures.

**Storage and Handling:** This CANNON® Certified Viscosity Reference Standard should be stored in the original container with the lid tightly closed, away from direct light, and at ambient temperatures and normal laboratory conditions. The standard was prepared in accordance with CANNON® Standard Laboratory Operating Procedures to ensure homogeneity and therefore mixing is unnecessary before use and no minimum sample volume is required.


**Composition and Product Safety:** This CANNON® Certified Viscosity Reference Standard is composed of: *Mineral Oil (100%)* [CAS#(s) 64742-54-7]. Consult MSDS for complete product safety information.

**Expiration of Certification:** The certification of this CANNON® Certified Viscosity Reference Standard is valid, within the stated measurement uncertainty, until the expiry date that appears on this certificate, provided the material is stored and handled as stated. This certification is deemed null and void if the standard is modified or contaminated. The shelf life was determined empirically through a historical evaluation of material stability. If substantive technical changes occur to the product which affects the certification before the expiry date, CANNON Instrument Company will contact the purchaser.

 <b>ISO/IEC 17025</b> ACCREDITED <small>Calibration Laboratory CERT#1262.01</small>	 <b>ISO Guide 34</b> ACCREDITED <small>Reference Material Producer CERT#1262.02</small>
<small>The inclusion of the A2LA and ILAC MRA logos does not imply certification/approval of the products calibrated or tested.</small>	
<b>ISO 9001</b> Registered by UL-DQS #10002540 QM	

BOTTLED BY  
IA

Certification Under Supervision of:

  
 \_\_\_\_\_  
 D.B. Trowbridge, Ph.D.  
 J.T. Mastropiero  
 M.T. Zubler

## DISCUSSION OF DATA<sup>1</sup>

**Derivation of Certified Values:** CANNON Instrument Company certifies that the kinematic viscosities were determined by the Master Viscometer technique reported in the Journal of Research of the National Bureau of Standards, (Vol. 52, No. 3, March 1954, Research Paper 2479) using CANNON® Laboratory Standard viscometers. All temperature measurements were conducted according to The International Temperature Scale of 1990 (ITS-90) using SPRTs with fixed point calibrations. The provided viscosity data are based upon the primary standard, water at 20 °C, with a kinematic viscosity of 1.0034 mm<sup>2</sup>/s and an assigned accuracy of ± 0.17% as per ISO 3666. See also ASTM methods D2162, D445, D446, D2161, and ISO methods 3104 and 3105.

Kinematic viscosity ( $\nu$ ) measurements in mm<sup>2</sup>/s at temperatures of 20, 25, 37.78, and 40 °C were generally made using CANNON® and/or Cannon-Ubbelohde (long capillary) Master viscometers, as described in ASTM methods D2162, D445, and D446. Measurements at other temperatures have been made using Cannon-Ubbelohde Laboratory Standard viscometers.

Density ( $\rho$ ) in g/cm<sup>3</sup> (g/mL) was generally determined through measurement in an oscillating U-tube digital density meter or modified Bingham pycnometer. See ASTM methods D4052, D1480, and D1217.

Dynamic viscosity ( $\eta$ ) in mPa·s was generally determined by measuring the kinematic viscosity and multiplying it by the density at the same temperature [ $\eta = \nu \cdot \rho$ ]. In some cases, dynamic viscosity was measured directly using Cannon-Manning Vacuum Laboratory Standard viscometers. See ASTM method D2171.

Where appropriate, the kinematic viscosity, dynamic viscosity, or density at certain temperatures was determined through regression of all measured data using industry standard equations. These equations include the linear or quadratic viscosity/density-temperature equation derived from the ASTM viscosity-temperature charts for petroleum products as well as the NBS viscosity-temperature equation for petroleum products. See ASTM method D341 and NBS equation.

Saybolt viscosity in Saybolt Universal Seconds (SUS) and in Saybolt Furol Seconds (SFS) was determined through mathematical conversion of measured kinematic viscosities in mm<sup>2</sup>/s. See ASTM method D2161.

**Traceability:** All data are traceable to intrinsic standards and National Institute of Standards and Technology (NIST) calibration or calculated by ASTM or NIST methods. Kinematic viscosity values are traceable to the viscosity of water. Temperature measurements were conducted with SPRTs that have NIST traceable fixed-point calibrations. A complete traceability statement is available for purchase from CANNON Instrument Company.

**Measurement Uncertainty:** CANNON Instrument Company has determined and reported the measurement uncertainty of its laboratory capabilities. The expanded uncertainties of the laboratory measurements summarized at the 95% confidence interval are as follows:

### Kinematic Viscosity (- 40 °C to + 150 °C)

Range of Kinematic Viscosity (mm <sup>2</sup> /s)	Expanded Uncertainty* (%) at Temperatures:		
	<15°C	15 to 45°C	>45°C
<10	0.21	0.16	0.21
10-100	0.26	0.22	0.26
100-1000	0.32	0.29	0.32
1000-10,000	0.47	0.38	0.38
10,000-100,000	0.53	0.44	0.48

### Density (- 56 °C to + 150 °C)

Range of Density (g/cm <sup>3</sup> )	Expanded Uncertainty* (kg/m <sup>3</sup> )
0.7 – 1.2	0.05

\* An expanded uncertainty  $U$  is determined by multiplying the combined standard uncertainty  $u_c$  by a coverage factor  $k$ :  $U = k u_c$  where  $k=2$ . See NIST Technical Note 1297, 1994 edition, Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results.

The expanded uncertainty for dynamic viscosity can be considered equivalent to the expanded uncertainty for kinematic viscosity since the uncertainty contribution of the density measurement is deemed negligible in the calculation of the total expanded uncertainty.

CANNON Instrument Company 2139 High Tech Road State College, PA 16803 USA 814 353-8000 • 800 676-6232 • FAX 814 353-8007	Laboratory Technical Director:	D.B. Trowbridge, Ph.D.
	Deputy Laboratory Technical Director:	J.T. Mastro pierro
	Director of Quality Assurance:	M.T. Zubler

<sup>1</sup>Consult [www.cannoninstrument.com](http://www.cannoninstrument.com) for additional information.